



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Element Materials Technology Food US LLC
12003 N.E. Ainsworth Circle, Suite 105, Portland, OR 97220

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Biological and Chemical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

June 5, 2014

Issue Date:

February 26, 2019

Expiration Date:

May 31, 2021

Revision Date:

March 17, 2020

Accreditation No.:

52833

Certificate No.:

L19-108-R2

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Element Materials Technology Food US LLC

12003 N.E. Ainsworth Circle, Suite 105, Portland, OR 97220

Contact Name: Leeza Akimenko Phone: 503-253-9136

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Biological ^F	Food products, Environmental, Water and Supplements	Aerobic Plate Count	FDA BAM Online Ed. Ch. 3; CMMEF 4 th Ed. Ch. 6.3 & 7.0 (Modified)	10 CFU/g (D.L. for solids)
		Aerobic Plate Count – Petrifilm	AOAC 2015.13	1 CFU/mL (D.L. for liquids)
		Yeast and Mold	FDA BAM Online Ed. Ch. 18; CMMEF 4 th Ed. Ch. 20.2; CMMEF 3 rd 16.52 (Modified)	
		Yeast and Mold Rapid – Petrifilm	AOAC RI 121301	
		Coliform - Petrifilm	AOAC 991.14	
		E. coli - Petrifilm		
		Enterobacteriaceae - Petrifilm	AOAC 2003.01	
		Staphylococcus aureus - Petrifilm	AOAC 2003.07	
		STEC (BAX PCR) (real-time)	AOAC RI 091301	Presumptive Positive / Negative
		Salmonella	AOAC R1 080601 (ELISA)	
			AOAC 2003.09 (PCR)	
			AOAC R1 051303 (PCR)	
		Listeria	AOAC RI 020401 (ELISA)	
			AOAC R1 071304 (PCR)	
	AOAC RI 030502 (PCR)			
	Listeria monocytogenes	AOAC 2003.12 (PCR)		
		AOAC R1 061302 (PCR)		
	E. coli O157:H7	AOAC RI 070801 (ELISA)		
		AOAC R1 021501 (PCR)		
		AOAC RI 050501 (PCR)		
Chemical ^F	Ash	AOAC 923.03	D.L. = 0.1 %	
	Fat	AOAC 960.39		
	Moisture	AOAC 925.10		
	Protein	AOAC 981.10		



Certificate of Accreditation: Supplement

Element Materials Technology Food US LLC

12003 N.E. Ainsworth Circle, Suite 105, Portland, OR 97220

Contact Name: Leeza Akimenko Phone: 503-253-9136

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical ^F	Food products, Environmental, Water and Supplements	Sodium Potassium Calcium Iron	EPA 7000B (modified)	D.L. = 1 mg/kg (Na) D.L. = 0.4 mg/kg (K) D.L. = 0.2 mg/kg (Ca) D.L. = 0.2 mg/kg (Fe)
		Arsenic Lead Cadmium Chromium	EPA 7010 (Modified)	D.L. = 0.001 mg/kg (As, Cr) D.L. = 0.0005 mg/kg (Pb, Cd)
		Sorbic Acid/Benzoic Acid	Internal HPLC Method SOP C250 rev. 5	D.L. = 0.1 mg/kg
		Vitamin C	Journal of Chromatography A, 881 (2000) 309-316 (modified)	D.L. = 1 mg/100g
		Water Activity	AOAC 978.18	0.021 aw to 1 aw
		pH	AOAC 981.12	1.68 pH to 12.46 pH

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this testing at its fixed location.